

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
<small>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.</small>				
1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE Jan 1998		3. REPORT TYPE AND DATES COVERED FINAL TECH RPT, 01 SEP 94 TO 31 AUG 97
4. TITLE AND SUBTITLE (AASERT 94) INTELLIGENT CONTROL OF MATERIALS PROCESSES				5. FUNDING NUMBERS F49620-94-1-0383
6. AUTHOR(S) Professor Patrick H. Garrett				AFRL-SR-BL-TR-98-0171
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Electrical and Computer Engineering Dept University of Cincinnati Cincinnati OH 45221				
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) AFOSR/NM 110 Duncan Avenue Suite B115 Bolling AFB DC 20332-8050				10. SPONSORING/MONITORING AGENCY REPORT NUMBER
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION AVAILABILITY STATEMENT Approved for public release; distribution unlimited.				12b. DISTRIBUTION CODE
13. ABSTRACT (Maximum 200 words) A systematic set of experiments has been performed to determine the effect of temperature, oxygen, precursor composition, and precursor flow rate on the gas phase of chemical vapor deposition.				
14. SUBJECT TERMS Keywords: precursor composition, vapor disposition				15. NUMBER OF PAGES 2
				16. PRICE CODE
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT UL	

Progress Report For A.F.O.S.R.

Progress has been made in the understanding, further refinement, and preventive maintenance of the chemical vapor deposition (CVD) system. A systematic set of experiments have been performed to determine the effect of temperature, oxygen, precursor composition, and precursor flow rate on the gas phase and *in situ* temperature of the process. Further refinement to the process was accomplished by adding two furnaces which can be heated up to 1200°C: one to desize the fiber before being coated, and a second furnace to serve as the reactor. This allowed the three zone furnace which was originally ineffectively used as a reactor to be used as an in-line heat treatment furnace at temperatures up to 1500°C.

As part of a preventive maintenance effort, a pneumatic valve was added to the sample gas inlet of the mass spectrometer. The pneumatic valve was interfaced to the computer system; and custom software was added to allow the operator to open or close the valve as needed. This prevents the sampling port from being open unnecessarily which could lead to permanent damage to the roughing pump if it happens over an extended period of time. An ion gauge and corresponding controller was added to the vacuum column to monitor the vacuum to prevent damage to the turbo molecular pump by operating it at an unacceptably poor vacuum. The vacuum is also interfaced to the computer as the mass spectrometer intensity is known to be dependent on pressure. Finally, a problem was detected when the turbo molecular pump would not start - after more than a week of troubleshooting it was correctly determined that two capacitors in the power supply were bad and promptly replaced. Fixing the problem prevented the shipment of the instrument back to the factory at a considerable expense and inconvenience.

Using the data acquired from the systematic set of experiments, a fuzzy logic controller was developed to regulate the *in situ* temperature of the process by varying the oxygen flow rate entering the reactor. Prior to actually regulating the temperature, a controller was developed which would use the mass spectrometer and regulate the gas phase with oxygen. These results are being published and being considered for filing of a patent.




Figure 1a. Molecular Materials Process Plan View